

HYDROPHILIC ACRYLIC FOLDABLE INTRAOCULAR LENS

STERILE

INSTRUCTIONS FOR USE



Amstermed BV

Saturnusstraat 46-62, Unit 032, 2132 HB Hoofddorp,
The Netherlands
email : info@amstermed.nl

INSTRUCTIONS FOR USE HYDROPHILIC ACRYLIC FOLDABLE INTRAOCULAR LENS

Intended Purpose

The Intraocular lenses (IOLs) are intended for the replacement of the human crystalline lens in the visual correction of the aphakia.

Device Description

Intraocular lens is composed of two elements: 1) Optic, and 2) Haptic. The optic is the small centre portion that acts as an artificial lens, and the haptic is the side structure that holds the lens in place in a capsular bag. It is implanted in the Posterior Chamber of the eye.

The raw materials used to manufacture of HYDROPHILIC intraocular lens is Contaflex 26% UV (C126CLSTUVRD) material is a Co polymer of Methyl Methacrylate and Poly Hydroxyl Ethyl Methacrylate + + UV Blocker AEHB -hydroxy-4-acryloxyethoxy benzophenone) the HYDROPHILIC IOL material have high water content, a lower refractive index, and increased IOL thickness.

A hydrophilic monofocal IOL is an intraocular lens with a fixed focus for one distance. The natural lens is protected against phototoxic UV light due to the filter properties of cornea, aqueous humor, lens and retina. After cataract surgery, it becomes important to protect the eye from UV radiation through artificial lens. The hydrophilic monofocal lens has a UV filter that allows transmission of <10% of light of wavelength 360 nm or lower which ensures protection of the eye from UV radiation. The Hydrophilic Monofocal IOLs are available in yellow optic tint, which is known to provide better contrast sensitivity by absorbing blue light and possibly protect the retina from damage.

Brand Name: G-FOLD

Variant/Model	Optic Dia	Overall Dia	Diopter
VFA 6010, VYA 6010	6.00 mm	11.00 mm	-5.0 D to +40.0 D
VFS 6025, VFS 6025D VFA 6025, VFA 6025D VYA 6025, VYA 6025D		12.50 mm	

Medical Indication

Cataract.

Intended User

Ophthalmic Surgeon

Target Patient Population

Above age of 18 years (Male or Female)

A Constant Information

The constant listed on the outer label is presented as a guideline and is a starting point for implant power calculation. It is recommended that surgeon calculates their own personalized A constant based on clinical experience with the particular IOL models, surgical techniques, measuring equipment and postoperative results.

Sterility

These posterior chamber intraocular lenses are supplied in a package terminally sterilized by using Steam Sterilization. Do not use the lens if the package is damaged. If used it leads to Inflammation.

Clinical benefits

- Excellent visual outcomes and a low incidence of PCO.
- Effective for the visual correction of aphakia.
- Excellent optical performance, good centration, and a low incidence of intra and postoperative complications.
- Visual improvement or stability in the majority of eyes (91.3%).
- Long-term safety, efficacy and rotational stability.
- Good uveal biocompatibility
- Low potential for endothelial damage
- Good resistance and low degree of adhesion
- Minimizing posterior capsule opacification (PCO)

Summary of Safety and Clinical Performance Referred

https://health.ec.europa.eu/system/files/2022-03/md_ndcg_2019_9_sscp_en.pdf

Recommendation for Choosing Lens Delivery System

The use of a lens delivery system is essential for the implantation of intraocular lens. It consists of the injector with silicone tip and a cartridge.

Directions For Use

- The Device is to be used by a qualified and trained ophthalmic surgeon only.
- The IOL should be maintain the temperature between 5°C (41°F) to 45°C (113°F) at the time of implantation.
- Information of the model, dioptre (power) and expiration date shown on the box label must be checked.
- Peel and open the pouch and remove the lens container under the sterile environment.
- The Dioptre of the IOL must be checked again from the label on the lens container.
- Open the lens container tray and peel the aluminum foil solely and take out the lens by forceps.
- Soak or rinse lens in sterile balanced salt solution (BSS) or their equivalent solution.
- Ensure that the lens is in good conditions for optic and haptics surfaces for adhesions of any particles.
- Before implantation examine the lens configuration data and optical surfaces
- Acrylic foldable IOLs are hydrated condition. They may be implanted either using a suitable injector system or a folding forceps.
- Packing box contains surgeons record label and patient ID card labels. These are for convenience in maintaining and reporting records of implantable lenses during clinical investigation.

Patient card:

The relevant details should be entered onto the patient card enclosed. One of the stickers with the IOL details from the label set enclosed should be affixed on the back of the patient card. This card is to be given to the patient, who should take care of it so as to present it to any eye specialist in the future.

Contraindications

Physicians considering lens implantation under any of the following circumstances should weigh the potential risk/benefit ratio.

Do not use in patients with pre-existing ocular conditions not explicitly listed below that may be aggravated by the implant or where the implant may interfere with the assessment or management of ocular disease.

- Progressive disturbances on the front segment of the eye such as rubeosis iridis, essential iris atrophy.
- Chronic or recurrent Uveitis.
- Proliferate diabetic retinopathy.
- Glaucoma Problem.
- Concomitant Severe Eye Disease.
- Severe optic nerve atrophy.
- Cataract associated with congenital rubella syndrome.
- Epithelial Systrophy.
- Rubella Cataract.
- Massive Vitreous loss.
- Microphthalmus.
- Do not use under age of 18.
- Patients in whom the intraocular lens may affect the ability to observe, diagnose or treat posterior segment diseases.

- A distorted eye due to previous trauma or developmental defect in which appropriate support of the IOL is not possible.

- Suspected microbial infection.

During surgery

Flat anterior chamber following clear lens extraction Circumstances that would result in damage to the endothelium during implantation. Surgical difficulties at the time of cataract extraction, which may increase the potential for complications (e.g., persistent bleeding, Hyphema, significant iris damage, retina detachment, uncontrolled positive pressure or significant vitreous prolapse or loss).

Pre-caution

Do not use the lens if the package has been damaged. The sterility of the lens may have been compromised.

Warnings

- Do not re-use the IOL. If used it can cause Inflammation.
- Do not autoclave/sterilized the Hydrophilic foldable intraocular lens.
- Do not soak or rinse the intraocular lens with any solution other than sterile balanced salt solution or sterile normal saline.
- Do not store the lens in direct sunlight.
- Storage packing solvent must not be used an eye irrigation solution.
- Never handle IOL by its optics.
- The IOL should not be allowed to get dry between being taken out of the blister and being applied to the eye and the foldable situation of the lens must be absolutely protected.

- Surgery must be performed using non-toothed, polished instruments.

- When IOL is handled prior to loading into the injector. Irrigate/aspirate to eliminate any viscoelastic residues from the bag, especially between the IOL and posterior capsule.

- Not to implant the IOL in the container which maintains sterility has been opened or damaged.

- The physician should have knowledge in selecting type of the lens depending on the eye dimensions.

Complications/Adverse Events

The Complications related to the implantation of any intraocular lens are essentially the same as for cataract surgery and may require secondary surgical intervention. Complications seen with the same type of IOL may include but are not limited to: Corneal Edema, Iritis, Hyphema, Macular Edema, Pupilary Block, Secondary Glaucoma, Cylitic Membrane, Vitritis, Endophthalmitis, Retinal Detachment, Lens Dislocation. Adverse events seen with same type of IOL may include but are not limited to: Hypopyon, Intraocular Infection, Acute Corneal Deкомпensation, Secondary Surgical Intervention: a) Lens replacement remover, b) Retinal Detachment Repair, c) Repositioning of lens, d) Iridectomy, e) Vitrectomy, f) Wound Repair Leak, g) Photocoagulation, h) Removal of Residual Cortex material, I) Anterior Capsulotomy.

Residual Risks

- Eye damage or loss of vision (May result in temporary or permanent vision loss or serious eye injury).

- Posterior Capsule Opacification (PCO) (Clouding of the lens capsule that may reduce vision), IOL dislocation (The implanted lens may shift from its intended position, affecting vision and possibly requiring surgical repositioning), Inflammation or infection (Endophthalmitis) (Postoperative inflammation or infection inside the eye may cause pain, redness, and vision impairment).

- Corneal Edema (Swelling of the cornea may cause blurred or hazy vision, which is usually temporary but can occasionally persist), Blurred Vision (Vision may be temporarily or permanently unclear following the procedure), Allergic Reaction (Rare allergic responses to lens materials or medications may cause ocular redness, swelling, or discomfort), Inconvenience to patient (night glares, halos) (Visual disturbances such as glare or halos may occur, particularly under low-light conditions).

- Reduced contrast sensitivity (trouble seeing things when they don't stand out well from the background).

Storage and Transport

Keep in cool and dry place. Store the Intraocular Lenses at temperatures between 5°C (41°F) to 45°C (113°F). Do not expose containers to direct sun light. Protect from freezing.

Material Used:

Contaflex 26% UV (C126CLSTUVRD) material is a Co polymer of Methyl Methacrylate and Poly Hydroxyl Ethyl Methacrylate + + UV Blocker AEHB -hydroxy-4-acryloxyethoxy benzophenone).

Disposal Condition:

Ensure safe and proper disposal of used/discarded product and packaging to avoid adverse effect to environment, children and stray animals. The disposal should be in compliance to local laws related to disposal of biomedical waste, in the country of use.

IOL Dioptic Power Calculations

The physician should determine preoperatively the power of the lens to be implanted Lens power calculation methods are described in the following references:

1. Hoffer, K.J. The Hoffer Q formula: A comparison of theoretic and regression formulas. Journal of Cataract and Refractive Surgery.1993;19:700-712; ERRATAL:1994;20:677.
2. Holladay, J.T., Musgrove, K.H., Prager, T.C., and Ruiz, R.S. A three-part system for refining intraocular lens power calculations. Journal of Cataract and Refractive Surgery. 1988; 14:17-24.
3. Holladay, J.T. Standardizing constants for ultrasonic biometry, keratometry and intraocular lens power calculations. Journal of Cataract and Refractive Surgery. 1997; 23:1356-1370.
4. Norrby NES. Unfortunate discrepancies. Letter to the editor and reply by Holladay, J.T. Journal of Cataract and Refractive Surgery. 1998; 24:433-434.

How Supplied

The Global Ophthalmic Acrylic Foldable IOLs are supplied sterile and hydrated in apyrogen saline solution enclosed in a sterile packaging and sterilized with steam Sterilization. The pouch and product labels are enclosed in a shelf pack.

Expiration Date

The expiration date on the lens package is the sterility expiration date. This lens should not be implanted after the indicated sterility expiration date.

Return of Damaged product

Return the lens in its original container identified by the batch number and serial number, model, purchase information, your reference and reason for the return. Please contact your local distribution office regarding lens return or exchange.

Reporting

Adverse events and/or potentially sight-threatening complications that may reasonably be regarded as lens related and that were not previously expected in nature, severity or incidence must be report to Global Ophthalmic Private Limited Company. Any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

Patient Information

Each patient should receive information regarding IOL prior to the decision to implant an intraocular lens.

Symbols

Symbol	Description	Symbol	Description
	Intraocular Lens		Date of manufacturing
	Ultraviolet		Use by (YYYY-MM: year - month)
	Posterior chamber		Authorized representative in the European Community
	Body diameter (Optic diameter)		Medical barrier system
	Overall diameter (Overall length)		Medical device
	Sterilized by using Steam		Catalogue Number
	Do not re-use		Serial Number
	Do not Sterilize		Country of Manufacturer
	Do not use if package is damaged		Manufacturer
	Caution		Model number
	Keep away from sunlight		Unique Device Identifier
	Storage between 5° to 45° C		Patient Identification
	Keep dry		Date
	Lot number / Batch number		Health Care Centre or Doctor
			Patient Information Website.

Mfg. Lic No.: MFG/MD/2019/000017

Manufactured by

Global Ophthalmic Pvt. Ltd.

Plot No.3, First Main Road., Sri Ramsamaji Nagar Extn.

Vellanoor Village, CHENNAI - 600062, Tamil Nadu, INDIA

Phone: +91-0-94456 01396, email: customercare@globalophthalmic.in